

Wojciech Kocon
Kleczewska 41
Warszawa
01-826
Poland

Jul 24, 2020

FMF ID: 74510

Preeclampsia screening - Measurement of uterine artery PI

Dear Wojciech Kocon,

Thank you for submitting your logbook of 3 images. I examined each image for the following criteria: sagittal section, angle of insonation angle less than 30 degrees, velocity more than 60 cm/s, and sample volume 2.0 mm (see table below)

Criteria	Image number		
	1	2	3
Sagittal section	✓	✓	✓
Insonation angle < 30°	✓	✓	✓
Velocity > 60 cm/s	✓	✓	✓
Sample Volume 2.0 mm	✓	✓	✓

Legend: ✓ = pass ✗ = fail -- = not possible to assess

Your images are satisfactory. Your software license will include the assessment of **uterine artery flow** in the risk calculation for preeclampsia and your name will appear on the list of sonographers certified for assessment of uterine artery flow.

Best wishes,

Ewelina Litwińska-Korcz