

Wojciech Kocon
Kleczewska 41
Warszawa
01-826
Poland

Jul 23, 2021

FMF ID: 74510

Preeclampsia screening - Measurement of uterine artery PI

Dear Wojciech Kocon,

Thank you for submitting your logbook of 3 images. I examined each image for the following criteria: sagittal section, angle of insonation angle less than 30 degrees, velocity more than 60 cm/s, and sample volume 2.0 mm (see table below)

| Criteria | Image number | | |
|------------------------|---------------------|----------|----------|
| | 1 | 2 | 3 |
| Sagittal section | ✓ | ✓ | ✓ |
| Insonation angle < 30° | ✓ | ✓ | ✓ |
| Velocity > 60 cm/s | ✓ | ✓ | ✓ |
| Sample Volume 2.0 mm | ✓ | ✓ | ✓ |

Legend: ✓ = pass ✗ = fail -- = not possible to assess

Your images are satisfactory. Your software license will include the assessment of **uterine artery flow** in the risk calculation for preeclampsia and your name will appear on the list of sonographers certified for assessment of uterine artery flow.

Best wishes,

Ewelina Litwińska-Korcz